

AMENDMENTS TO THE SPECIFICATION

Page 6, please amend the third full paragraph as follows:

The interchangeable attachment means 4 is held in place in the aperture 17 by means of a carrier part 19 of the hearing aid, carrying an output transducer 20 and preferably all or parts of the electronics of the hearing aid. The interchangeable attachment part 4 comprises a number of radial protrusions 21. When the hearing aid is assembled, the carrier part 19 within the hearing aid housing 1, presses the radial protrusions ~~19~~21 of the attachment part 4 against the internal side of the end wall 18, thereby securing the attachment part 4 against linear motion. In order to also secure the attachment part 4 against rotation with respect to the hearing aid housing 1, the carrier part 19 has a number of cut outs ~~21~~ matching the protrusions ~~20~~21. In this respect, it should be noted that with appropriate choice of the width of the cut outs ~~19~~, a flange plate corresponding to or even identical to that of Figs. 3 and 4 could be used instead of the protrusions ~~20~~21.